



UNIVERSITÀ POLITECNICA DELLE MARCHE
Repository ISTITUZIONALE

Characterization of Analog-to-Information Converters

This is the peer reviewed version of the following article:

Original

Characterization of Analog-to-Information Converters / Iadarola, G. - In: IEEE INSTRUMENTATION & MEASUREMENT MAGAZINE. - ISSN 1094-6969. - 25:1(2022), pp. 98-99. [10.1109/MIM.2022.9693447]

Availability:

This version is available at: 11566/310992 since: 2024-05-17T17:29:08Z

Publisher:

Published

DOI:10.1109/MIM.2022.9693447

Terms of use:

The terms and conditions for the reuse of this version of the manuscript are specified in the publishing policy. The use of copyrighted works requires the consent of the rights' holder (author or publisher). Works made available under a Creative Commons license or a Publisher's custom-made license can be used according to the terms and conditions contained therein. See editor's website for further information and terms and conditions.

This item was downloaded from IRIS Università Politecnica delle Marche (<https://iris.univpm.it>). When citing, please refer to the published version.

(Article begins on next page)

© 2022 IEEE. Personal use of this material is permitted. Permission from IEEE must be obtained for all other uses, in any current or future media, including reprinting/republishing this material for advertising or promotional purposes, creating new collective works, for resale or redistribution to servers or lists, or reuse of any copyrighted component of this work in other works.

Characterization of Analog-to-Information Converters

The dissertation deals with the characterization of sub-sampling devices, the Analog-to-Information Converters (AICs), recently proposed in research as an alternative to the traditional Analog-to-Digital Converters. By exploiting the Compressive Sampling paradigm, AICs stably recover the input signal, previously acquired with less samples than needed by the Nyquist-Shannon theorem, reducing simultaneously sampling frequency and acquisition memory. The dissertation research was carried out in order to design and develop an AIC prototype to be employed in a measurement instrument: a vector signal analyzer, that acquires and reconstructs sparse wideband signals. The aim of such an instrument is to guarantee proper monitoring, circumventing the trade-off between resolution and sampling frequency, typical of conventional data acquisition systems. With a view to this use, the characterization of AICs is of crucial importance, since their performance conditions the measurement instrument based on them, and, although several efforts have been made in research to propose several architectures, only in a few cases their performance is assessed.

Firstly, a state of the art on AICs is presented in the dissertation, by classifying the existing architectures, prototypes and models. Then, the dissertation focuses on the characterization of AICs, by identifying non-idealities affecting the individual components and by recommending testing methods and corresponding figures of merit for AIC assessment, differentiating them in methods before and after the reconstruction of the input signal. Specifically, a new figure of merit adopted before the reconstruction is proposed. Thus, the experimental contribution of the dissertation is presented: the analysis of AIC performance, depending on the non-idealities through the aforementioned figures of merit. In particular, two categories of AICs are considered for the metrological characterization: (i) the Random Demodulation (RD) AIC and the similar counterpart structured on parallel channels, the Modulated Wideband Converter (MWC); (ii) the Non-Uniform Wavelet Bandpass Sampling (NUWBS) AIC. With regard to the first AIC category, several investigations were performed: the comparison between two mathematical models for the same prototype and the analysis related to non-idealities of the two building blocks involved in the modulation of the input

signal, i.e. pseudo-random sequence and mixer. With regard to the second category, several investigations were performed as well: the analysis related to main noise sources and the experimental assessment of a proposed prototype employing a low pass-filter instead of an analog integrator.

Based on the results of the dissertation (Fig. 1), on one hand, the experimental analysis suggests the use of specific components rather than others to implement the AIC board for the vector signal analyzer, namely an analog multiplier instead of a mixer, to reduce the distortion phenomenon, or a Wavelet signal, if the monitored frequency interval of the input signal requires to be extended. On the other hand, the numerical analysis suggests key aspects on the measurement matrix to be adopted in the reconstruction phase, such as the importance of including architectural non-idealities and actual features of the prototype, first of all the pulse shape of signals involved in modulation processes and the impulse/frequency response of filters located ahead ADCs working at sub-Nyquist rates.




Fig. 1. Phases for characterization of AICs.

References

- [1] G. Iadarola, "Characterization of Analog-to-Information Converters - Testing methods for analysis of non-idealities," *Ph.D. dissertation*, Dept. Eng., Univ. of Sannio, Benevento, Italy, 2019.

- [2] L. De Vito, G. Iadarola *et al.*, “Non-uniform wavelet bandpass sampling analog-to-information converter: a hardware implementation and its experimental assessment,” *Measurement*, vol. 134, pp. 739–749, Feb. 2019.
- [3] P. Daponte, L. De Vito, G. Iadarola, S. Rapuano, “A reduced-code method for integral nonlinearity testing in DACs,” *Measurement*, vol. 182, Sep. 2021.
- [4] P. Daponte, L. De Vito, G. Iadarola, and S. Rapuano, “Effects of PRBS jitter on random demodulation analog-to-information converters,” in *Proc. IEEE Int. Workshop on Metrology for Aerospace*, pp. 630–635, 2016.
- [5] P. Daponte, L. De Vito, G. Iadarola *et al.*, “Experimental comparison of two mathematical models for Analog-to-Information Converter,” in *Proc. 21st IMEKO TC4 Int. Symp. on Understanding the World through Elect. and Electron. Meas., and 19th Int. Workshop on ADC Model. and Testing*, pp. 65–70, 2016.
- [6] P. Daponte, L. De Vito, G. Iadarola, and S. Rapuano, “PRBS non-idealities affecting random demodulation analog-to-information converters,” in *Proc. 21st IMEKO TC4 Int. Symp. on Understanding the World through Elect. and Electron. Meas., and 19th Int. Workshop on ADC Model. and Testing*, pp. 71–76, 2016.
- [7] P. Daponte, L. De Vito, G. Iadarola, and S. Rapuano, “Experimental characterization of a RF mixer for wideband data acquisition systems,” in *Proc. IEEE Int. Instrum. and Meas. Technol. Conf.*, 2017.
- [8] P. Daponte, L. De Vito, G. Iadarola, and S. Rapuano, “Analog multiplication in random demodulation analog-to-information converters,” *J. Physics: Conf. Series*, vol. 1065, no. 5, Nov. 2018.
- [9] P. Daponte, L. De Vito, G. Iadarola, and S. Rapuano, “A CS method for DAC nonlinearity testing,” in *Proc. 24th IMEKO TC4 Int. Symp. and 22nd Int. Workshop on ADC and DAC Model. and Testing*, pp. 146–150, 2020.
- [10] P. Daponte, L. De Vito, G. Iadarola, and S. Spinsante, “PRBS selection for velocity measurements with compressive sampling-based DS-CDMA radio navigation receivers,” in *Proc. IEEE Int. Workshop on Metrology for Aerospace*, pp. 305–310, 2018.

Grazia Iadarola (M’16) (g.iadarola@staff.univpm.it) is a Post-Doctoral Fellow in the field of instrumentation and measurement at the Polytechnic University of Marche, Italy. She received the Ph.D. degree in information technologies for engineering in 2019 from the University of Sannio, Italy, where she conducted post-doctoral research until 2021. She earned cum laude both her Bachelor’s degree in telecommunications engineering and her

Master's degree in electronic engineering from the University of Sannio, in 2013 and in 2015, respectively. Her research interests include sub-Nyquist sampling, characterization and testing of data converters, modeling of electronic circuits and non-idealities, signal reconstruction based on compressed sensing, as well as their applications to telecommunications and biomedical instrumentation.